

Document name: Certificate of Conformity		Issue date: 08-MAR-2021
Form F51-Certificate of Conformity		Revision 005

CERTIFICATE OF CONFORMITY

Product 1694, IUI Catheter soft short

Lot 210199

Sterilization The products are sterilized by gamma irradiation.

Intended use The products are intended for use in IVF/ART procedures.

The product/lot complies with our minimum quality requirements such as:

MEA $\geq 80\%$ embryo development to blastocyst stage

LAL <20 EU/Device

We declare in own responsibility, that the referenced products meet the General requirements of the guideline 93/42/EEC/Annex I.

All supporting documentation is retained at the premises of the Manufacturer.

Sterility is guaranteed unless the peel pouch is damaged or the product expiry date has expired.

Manufacturer Birr BioSciences B.V.
Bergseweg 4
3633 AK Vreeland, the Netherlands

CE Certificate CE 575680 BSI

Issue date CoC 11-05-2021



Loïc Brandsma
QA&RA Manager