

Document name: Certificate of Conformity		Issue date: 08-MAR-2021
Form F51-Certificate of Conformity		Revision 005

CERTIFICATE OF CONFORMITY

Product 113006, 6-Well Dish per 6

Lot 21327

Sterilization The products are sterilized by gamma irradiation.

Intended use IVF/ART

The product/lot complies with our minimum quality requirements such as:

MEA ≥ 80% embryo development to blastocyst stage

LAL <20 EU/Device

SMA ≥ 0.75

We declare in own responsibility, that the referenced products meet the General requirements of the guideline 93/42/EEC/Annex I.

All supporting documentation is retained at the premises of the Manufacturer.

Sterility is guaranteed unless the peel pouch is damaged or the product expiry date has expired.

Manufacturer Birr BioSciences B.V.
Bergseweg 4
3633 AK Vreeland, the Netherlands

CE Certificate CE 575680 BSI

Issue date CoC 12-08-2021



Loïc Brandsma
QA&RA Manager