

|   |   |                         |
|---|---|-------------------------|
| Document name:<br>Certificate of Conformity |  | Issue date: 08-MAR-2021 |
| Form F51-Certificate of Conformity          |   | Revision 005            |

## CERTIFICATE OF CONFORMITY

---

Product                    113411, Dropping Pipette 1 ml per 10

Lot                         210960

Sterilization             The products are sterilized by gamma irradiation.

Intended use             IVF/ART

The product/lot complies with our minimum quality requirements such as:

MEA                        ≥ 80% embryo development to blastocyst stage

LAL                         <20 EU/Device

SMA                        ≥ 0.75

We declare in own responsibility, that the referenced products meet the General requirements of the guideline 93/42/EEC/Annex I.

All supporting documentation is retained at the premises of the Manufacturer.

Sterility is guaranteed unless the peel pouch is damaged or the product expiry date has expired.

Manufacturer             Birr BioSciences B.V.  
Bergseweg 4  
3633 AK Vreeland, the Netherlands

CE Certificate            CE 575680 BSI

Issue date CoC           03-08-2021



Loïc Brandsma  
QA&RA Manager