

Document name: Certificate of Conformity		Issue date: 11-OCT-2019
Form F51-Certificate of Conformity		Revision 004

## CERTIFICATE OF CONFORMITY

---

Product                    1134+8, 4+8 Well Dish per 10

Lot                         20232

Sterilization            The products are sterilized by gamma irradiation.

Intended use            IVF/ART

The product/lot complies with our minimum quality requirements such as:

MEA                         $\geq 80\%$  embryo development to blastocyst stage

LAL                         $<20$  EU/Device

SMA                         $\geq 0.75$

We declare in own responsibility, that the referenced products meet the General requirements of the guideline 93/42/EEC/Annex I.

All supporting documentation is retained at the premises of the Manufacturer.

Sterility is guaranteed unless the peel pouch is damaged or the product expiry date has expired.

Manufacturer            Birr BioSciences B.V.  
Bergseweg 4  
3633 AK Vreeland, the Netherlands

CE Certificate           CE 575680 BSI

Issue date CoC        26-10-2020

Greta Ditzel i.o.  
Addy van der Klis  
QA&RA Manager