

CERTIFICATE OF ANALYSIS

Customer Name	BIRR BIOSCIENCES B.V.
Product Description	Disposable Needle Guide
Product Code / Size	BB-GTK08 / 16G
Invoice No. and Date	494865 / 11.12.2017
Lot #	#01080712

CHEMICAL COMPOSITION:

Cannula	C	Mn	Si	S	P	Cr.	Ni
Standard Max.	0.080	2.00	0.75	0.03	0.04	18-20.0	8-11.0
Material	0.0199	1.148	0.65	0.0023	0.0260	18.28	9.34

Body	ABS
Adhesive	Vitralit 6108

Pull Strength: Passed 50N
Finish & Look: Clean & Clear
Packing : Individual Packed & Sterile & Tyvek Pouches

WE CERTIFY THAT ALL MATERIAL TO BE TYPE 304 STAINLES STEEL WHERE THE CHEMICAL COMPOSITION ADHERES TO ASTM A908-03. WE CERTIFY THAT ALL PRODUCTS HAVE BEEN PROCESSED AND AUDITED TO ENSURE COMPLIANCE TO GEOTEK QUALITY SYSTEM REQUIREMENTS AS PER QP-04 "PROCESS CONTROL PROCEDURE" AND TRACEABILITY IS AVAILABLE UPON REQUEST.

Onursal Ersan
Product Manager

