

### CERTIFICATE OF ANALYSIS

Customer Name	BIRR BIOSCIENCES B.V.
Product Description	Disposable Needle Guide
Product Code / Size	BB-GTK141 / 16G
Invoice No. and Date	BRR180509 / 09.05.2018
Lot #	#011410705

#### CHEMICAL COMPOSITION:

<b>Cannula</b>	C	Mn	Si	S	P	Cr.	Ni
Standard Max.	0.080	2.00	0.75	0.03	0.04	18-20.0	8-11.0
Material	0.0199	1.148	0.65	0.0023	0.0260	18.28	9.34

<b>Body</b>	ABS
<b>Adhesive</b>	Vitralit 6108

**Pull Strength:** Passed 50N  
**Finish & Look:** Clean & Clear  
**Packing :** Individual Packed & Sterile & Tyvek Pouches

WE CERTIFY THAT ALL MATERIAL TO BE TYPE 304 STAINLES STEEL WHERE THE CHEMICAL COMPOSITION ADHERES TO ASTM A908-03. WE CERTIFY THAT ALL PRODUCTS HAVE BEEN PROCESSED AND AUDITED TO ENSURE COMPLIANCE TO GEOTEK QUALITY SYSTEM REQUIREMENTS AS PER QP-04 "PROCESS CONTROL PROCEDURE" AND TRACEABILITY IS AVAILABLE UPON REQUEST.

Onursal Ersan  
Product Manager

